
Midmark M11 - v1.0.5
Total Cycles: 1197

Name: RITTER2
Sterilizer ID: V2313228
06/16/2022 12:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 165.7 F 2.6
4:00 183.2 F 6.7
6:00 199.4 F 12.1
8:00 224.0 F 7.5
10:00 236.9 F 12.8
12:00 237.3 F 10.3
14:00 236.1 F 8.4
16:00 248.2 F 15.1
18:00 260.0 F 21.7
20:00 271.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.0
1:00 272.1 F 29.9
2:00 272.4 F 29.9
3:00 272.8 F 30.0
4:00 272.8 F 30.2

Min 271.7 F 29.9
Max 273.3 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 233.5 F 0.0
10:00 221.9 F 0.0
15:00 219.9 F 0.0
20:00 213.9 F 0.0
25:00 215.0 F 0.0
30:00 212.1 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

06/16/2022 01:40 PM
ACCEPTED: X
REJECTED:
Operator: U02