
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007461
10:58 29/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:19
STERILIZING 4:17
136.0 C 220kPa 4:17
135.9 C 221kPa 4:47
136.4 C 218kPa 5:17
136.2 C 221kPa 5:47
136.4 C 224kPa 6:17
136.4 C 223kPa 6:47
136.1 C 222kPa 7:17
136.3 C 223kPa 7:47
136.3 C 223kPa 7:47
VENTING 7:47
AIR DRYING 8:39
CASSETTE REMOVED
CYCLE COMPLETE 18:30

Digital Signature #
D6622C8A7A47B43B
