
Midmark M11 - v1.0.5
Total Cycles: 3376

Name: RITTER2
Sterilizer ID: V2313228
02/12/2024 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 167.4 F 1.7
4:00 182.2 F 5.6
6:00 204.1 F 10.9
8:00 222.0 F 5.9
10:00 234.0 F 11.4
12:00 235.4 F 9.1
14:00 238.4 F 9.7
16:00 249.0 F 15.1
18:00 261.8 F 22.3
20:00 272.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.2 F 29.1
2:00 271.7 F 29.3
3:00 272.1 F 29.0
4:00 271.7 F 29.3

Min 271.1 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 233.7 F 0.0
10:00 194.9 F 0.0
15:00 189.6 F 0.0
20:00 186.1 F 0.0
25:00 183.8 F 0.0
30:00 182.2 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

02/12/2024 12:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____