
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007668
7:46 15/07/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:04
PRESSURIZING 4:04
STERILIZING 5:09
136.4 C 222kPa 5:09
136.4 C 220kPa 5:39
136.2 C 219kPa 6:09
136.3 C 223kPa 6:39
136.3 C 222kPa 7:09
136.4 C 220kPa 7:39
136.4 C 222kPa 8:09
136.1 C 221kPa 8:39
136.1 C 221kPa 8:39
VENTING 8:39
AIR DRYING 9:30
CYCLE COMPLETE 11:12

Digital Signature #
F5EC75671D837754
