

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4402

Name: RITTER2  
Sterilizer ID: V2313228  
03/05/2025 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.1 F 0.0  
2:00 153.2 F 2.2  
4:00 167.3 F 5.6  
6:00 192.0 F 9.7  
8:00 207.3 F 10.0  
10:00 220.8 F 9.0  
12:00 231.9 F 13.5  
14:00 234.9 F 10.5  
16:00 234.3 F 8.4  
18:00 246.2 F 14.5  
20:00 257.4 F 20.5  
22:00 267.5 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 30.3  
1:00 272.0 F 30.6  
2:00 271.5 F 30.2  
3:00 271.8 F 30.6  
4:00 271.5 F 30.4

Min 271.4 F 30.2  
Max 272.6 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 199.4 F 0.0  
10:00 203.2 F 0.0  
15:00 205.0 F 0.0  
20:00 206.8 F 0.0  
25:00 208.8 F 0.0  
30:00 208.5 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 23:17  
STERILIZING: 4:00  
VENTING: 3:58  
DRYING: 30:00  
TOTAL CYCLE: 01:02:37  
CYCLE COMPLETE

03/05/2025 03:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_