
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002116
10:46 03/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:07
STERILIZING 3:13
136.4 C 223kPa 3:13
136.1 C 220kPa 3:43
136.4 C 226kPa 4:13
136.4 C 226kPa 4:43
136.4 C 224kPa 5:13
136.0 C 222kPa 5:43
136.4 C 223kPa 6:13
136.2 C 224kPa 6:43
136.2 C 224kPa 6:43
VENTING 6:43
AIR DRYING 7:34
CYCLE COMPLETE 37:34

Digital Signature #
67D30ED35F235F16
