
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000803
11:01 04/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:16
STERILIZING 4:32
136.4 C 233kPa 4:32
136.3 C 232kPa 5:02
136.4 C 233kPa 5:32
136.4 C 231kPa 6:02
136.4 C 233kPa 6:32
136.4 C 233kPa 7:02
136.4 C 233kPa 7:32
136.3 C 233kPa 8:02
136.3 C 233kPa 8:02
VENTING 8:02
AIR DRYING 8:57
CYCLE COMPLETE 8:58

Digital Signature #
BBB5F37A298BC7DA
