

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001582  
11:48 11/15/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:26  
PRESSURIZING 6:57  
STERILIZING 9:02  
136.4 C 220kPa 9:02  
136.5 C 222kPa 9:32  
136.3 C 221kPa 10:02  
136.3 C 219kPa 10:32  
135.7 C 217kPa 11:02  
136.4 C 221kPa 11:32  
136.4 C 217kPa 12:02  
136.4 C 221kPa 12:32  
VENTING 12:32  
AIR DRYING 13:38  
CYCLE COMPLETE 36:06

-----  
Digital Signature #  
F73E3897B92FD94A  
-----