

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005266  
8:39 27/03/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:23  
STERILIZING 3:27  
136.0 C 220kPa 3:27  
136.3 C 222kPa 3:57  
136.3 C 222kPa 4:27  
136.4 C 223kPa 4:57  
136.3 C 220kPa 5:27  
136.4 C 222kPa 5:57  
136.4 C 223kPa 6:27  
136.4 C 220kPa 6:57  
136.4 C 220kPa 6:57  
VENTING 6:57  
AIR DRYING 7:46  
CYCLE COMPLETE 7:50

-----  
Digital Signature #  
38975158A6AA758F  
-----