
Midmark M11 - v1.0.0
Total Cycles: 2676

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2022 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.2 F 0.0
2:00 167.4 F 3.3
4:00 185.0 F 7.3
6:00 205.1 F 12.3
8:00 218.0 F 6.5
10:00 231.8 F 11.7
12:00 234.0 F 8.6
14:00 239.9 F 10.2
16:00 246.8 F 13.6
18:00 259.1 F 20.4
20:00 270.5 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.0
2:00 271.5 F 28.3
3:00 272.5 F 29.1
4:00 271.9 F 28.5

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 230.3 F 0.0
10:00 234.9 F 0.0
15:00 228.2 F 0.0
20:00 225.5 F 0.0
25:00 229.5 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

12/14/2022 02:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____