
Midmark M11 - v1.0.0
Total Cycles: 626

Name: MIDMARK
Sterilizer ID: V2178198
01/25/2021 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.9 F 0.0
2:00 120.4 F 0.4
4:00 135.5 F 2.2
6:00 169.5 F 5.8
8:00 194.3 F 10.6
10:00 212.6 F 4.4
12:00 225.1 F 9.6
14:00 234.3 F 8.6
16:00 241.9 F 12.5
18:00 240.5 F 10.9
20:00 254.4 F 17.9
22:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.0 F 28.9
2:00 272.2 F 29.4
3:00 271.3 F 28.5
4:00 272.4 F 29.2

Min 271.1 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 235.5 F 0.0
10:00 224.0 F 0.0
15:00 215.4 F 0.0
20:00 211.9 F 0.0
25:00 210.3 F 0.0
30:00 209.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:27
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

01/25/2021 05:01 PM
ACCEPTED: X
REJECTED:
Operator: U01