
Midmark M11 - v1.0.5
Total Cycles: 636

Name: RITTER2
Sterilizer ID: V2313228
12/15/2021 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.9 F 0.0
2:00 170.3 F 3.1
4:00 178.3 F 7.3
6:00 201.5 F 12.6
8:00 218.7 F 7.1
10:00 232.8 F 12.7
12:00 236.1 F 10.2
14:00 235.4 F 8.7
16:00 250.2 F 15.9
18:00 262.2 F 23.0
20:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.4 F 29.8
2:00 271.5 F 29.1
3:00 271.8 F 29.9
4:00 272.4 F 29.7

Min 271.1 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 200.0 F 0.0
10:00 214.9 F 0.0
15:00 220.9 F 0.0
20:00 224.7 F 0.0
25:00 227.0 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

12/15/2021 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____