
Midmark M11 - v1.0.0
Total Cycles: 4548

Name: RITTER 1
Sterilizer ID: V2178198
05/13/2024 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.3 F 0.0
2:00 94.1 F 0.0
4:00 104.4 F 0.8
6:00 135.9 F 3.2
8:00 167.4 F 5.8
10:00 184.8 F 9.1
12:00 197.3 F 12.6
14:00 211.0 F 4.4
16:00 220.8 F 7.9
18:00 231.1 F 11.8
20:00 230.0 F 7.7
22:00 242.1 F 12.7
24:00 239.5 F 10.5
26:00 250.9 F 16.0
28:00 260.9 F 21.6
30:00 270.3 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.2
2:00 271.9 F 28.7
3:00 271.2 F 28.5
4:00 272.3 F 29.3

Min 271.2 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 203.4 F 0.0
10:00 196.8 F 0.0
15:00 189.3 F 0.0
20:00 211.7 F 0.0
25:00 225.1 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:58
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:09:43
CYCLE COMPLETE

05/13/2024 04:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____