
Midmark M11 - v1.0.0
Total Cycles: 4032

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2023 03:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.4 F 0.0
2:00 174.6 F 3.6
4:00 183.7 F 7.9
6:00 209.9 F 12.9
8:00 222.7 F 7.2
10:00 234.5 F 12.2
12:00 234.4 F 9.2
14:00 240.6 F 9.8
16:00 247.1 F 13.8
18:00 259.2 F 20.4
20:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.7 F 29.2
2:00 271.4 F 28.5
3:00 272.2 F 29.3
4:00 272.4 F 28.9

Min 271.2 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 228.2 F 0.0
10:00 235.3 F 0.0
15:00 227.0 F 0.0
20:00 224.6 F 0.0
25:00 227.5 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:56
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

11/16/2023 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____