

-----  
Midmark M11 - v1.0.5  
Total Cycles: 410

Name: RITTER2  
Sterilizer ID: V2313228  
09/30/2021 10:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.1 F 0.0  
2:00 173.4 F 3.1  
4:00 186.8 F 7.5  
6:00 208.1 F 13.4  
8:00 225.3 F 9.3  
10:00 231.8 F 9.0  
12:00 243.0 F 13.3  
14:00 243.8 F 12.5  
16:00 256.9 F 19.5  
18:00 268.5 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 271.5 F 29.0  
2:00 271.5 F 28.9  
3:00 271.8 F 29.1  
4:00 272.1 F 29.4  
  
Min 271.3 F 28.9  
Max 272.4 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.4 F 0.0  
5:00 228.0 F 0.0  
10:00 225.0 F 0.0  
15:00 225.4 F 0.0  
20:00 230.7 F 0.0  
25:00 228.6 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

09/30/2021 11:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02