

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3988

Name: RITTER2  
Sterilizer ID: V2313228  
10/02/2024 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.9 F 0.0  
2:00 156.0 F 2.0  
4:00 170.1 F 5.1  
6:00 188.2 F 8.9  
8:00 206.5 F 12.9  
10:00 221.2 F 6.7  
12:00 228.9 F 11.0  
14:00 230.2 F 7.4  
16:00 239.5 F 12.7  
18:00 240.4 F 11.0  
20:00 252.0 F 16.6  
22:00 261.7 F 22.5  
24:00 271.0 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.3 F 29.4  
2:00 272.2 F 29.8  
3:00 271.9 F 29.8  
4:00 271.6 F 29.3

Min 271.5 F 29.3  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 199.5 F 0.0  
10:00 194.3 F 0.0  
15:00 191.4 F 0.0  
20:00 187.8 F 0.0  
25:00 184.2 F 0.0  
30:00 183.8 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 25:45  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:04:50  
CYCLE COMPLETE

10/02/2024 12:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_