
Midmark M11 - v1.0.5
Total Cycles: 714

Name: RITTER2
Sterilizer ID: V2313228
01/20/2022 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.6 F 0.0
2:00 193.3 F 4.4
4:00 215.1 F 10.8
6:00 226.2 F 7.4
8:00 239.8 F 12.8
10:00 242.3 F 12.5
12:00 243.6 F 12.3
14:00 257.2 F 19.6
16:00 269.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.4
2:00 271.5 F 28.9
3:00 272.6 F 29.5
4:00 271.8 F 28.8

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 213.1 F 0.0
10:00 227.7 F 0.0
15:00 228.0 F 0.0
20:00 227.3 F 0.0
25:00 228.4 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 2:51
HEATING: 17:03
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:56:60
CYCLE COMPLETE

01/20/2022 01:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____