
Midmark M11 - v1.0.0
Total Cycles: 3481

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2023 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.4 F 0.0
2:00 166.3 F 2.7
4:00 178.9 F 6.4
6:00 204.4 F 11.3
8:00 221.1 F 5.5
10:00 230.8 F 10.6
12:00 233.7 F 7.2
14:00 243.7 F 13.1
16:00 242.0 F 11.5
18:00 254.8 F 18.0
20:00 265.9 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.2
2:00 271.7 F 28.4
3:00 272.1 F 29.1
4:00 272.3 F 29.0

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 226.9 F 0.0
10:00 229.8 F 0.0
15:00 228.9 F 0.0
20:00 228.6 F 0.0
25:00 230.2 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

07/04/2023 02:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____