
Midmark M11 - v1.0.5
Total Cycles: 3201

Name: RITTER2
Sterilizer ID: V2313228
12/11/2023 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 80.1 F 0.0
4:00 105.9 F 0.2
6:00 102.4 F 1.5
8:00 137.2 F 4.2
10:00 164.4 F 7.1
12:00 199.2 F 10.7
14:00 213.4 F 4.4
16:00 225.8 F 7.5
18:00 232.5 F 11.4
20:00 229.0 F 7.1
22:00 241.2 F 12.4
24:00 238.9 F 10.4
26:00 251.4 F 16.6
28:00 262.2 F 22.9
30:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.4
2:00 272.1 F 29.7
3:00 271.8 F 29.1
4:00 272.6 F 29.8

Min 271.2 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 232.6 F 0.0
10:00 212.0 F 0.0
15:00 201.4 F 0.0
20:00 195.3 F 0.0
25:00 233.3 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:30
CYCLE COMPLETE

12/11/2023 10:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____