
Midmark M11 - v1.0.5
Total Cycles: 1895

Name: RITTER2
Sterilizer ID: V2313228
01/04/2023 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 174.3 F 3.4
4:00 189.9 F 8.0
6:00 209.8 F 12.2
8:00 225.4 F 9.0
10:00 235.8 F 9.9
12:00 242.3 F 13.2
14:00 243.5 F 12.7
16:00 257.1 F 19.9
18:00 269.2 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.2 F 29.7
2:00 271.6 F 29.5
3:00 272.0 F 29.4
4:00 271.9 F 29.8

Min 271.4 F 29.1
Max 272.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 235.8 F 0.0
10:00 234.3 F 0.0
15:00 223.2 F 0.0
20:00 217.7 F 0.0
25:00 230.0 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

01/04/2023 04:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____