

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004394  
11:51 20/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:09  
PRESSURIZING 2:11  
STERILIZING 3:11  
136.0 C 220kPa 3:11  
136.4 C 225kPa 3:41  
136.4 C 225kPa 4:11  
136.3 C 221kPa 4:41  
136.3 C 224kPa 5:11  
136.4 C 224kPa 5:41  
136.3 C 225kPa 6:11  
136.2 C 219kPa 6:41  
136.2 C 219kPa 6:41  
VENTING 6:41  
AIR DRYING 7:31  
CYCLE COMPLETE 7:51  
-----

Digital Signature #  
49F360A8C9451EB9  
-----