

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4797

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2024 01:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.5 F 0.0  
2:00 151.0 F 0.7  
4:00 175.7 F 3.3  
6:00 194.8 F 7.1  
8:00 214.5 F 11.4  
10:00 221.9 F 4.9  
12:00 233.8 F 9.5  
14:00 238.9 F 10.8  
16:00 241.8 F 11.3  
18:00 238.1 F 9.5  
20:00 251.4 F 15.8  
22:00 262.4 F 22.1  
24:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.3 F 28.7  
2:00 271.3 F 28.0  
3:00 272.1 F 28.8  
4:00 272.0 F 28.5

Min 271.1 F 28.0  
Max 272.4 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
5:00 217.0 F 0.0  
10:00 196.7 F 0.0  
15:00 197.5 F 0.0  
20:00 206.8 F 0.0  
25:00 228.9 F 0.0  
30:00 231.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 24:25  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:03:04  
CYCLE COMPLETE

08/08/2024 02:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_