
Midmark M11 - v1.0.5
Total Cycles: 1719

Name: RITTER2
Sterilizer ID: V2313228
11/14/2022 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.8 F 0.0
2:00 184.3 F 2.0
4:00 200.7 F 7.3
6:00 220.1 F 13.1
8:00 229.3 F 9.2
10:00 236.2 F 9.5
12:00 243.3 F 12.0
14:00 246.7 F 13.9
16:00 259.9 F 21.2
18:00 271.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.3 F 29.1
2:00 272.3 F 29.5
3:00 272.0 F 28.8
4:00 272.5 F 29.6

Min 271.6 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 224.1 F 0.0
10:00 236.4 F 0.0
15:00 229.4 F 0.0
20:00 225.9 F 0.0
25:00 227.1 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:27
CYCLE COMPLETE

11/14/2022 04:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____