
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001450
10:44 09/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 6:10
STERILIZING 7:42
136.0 C 220kPa 7:42
136.4 C 223kPa 8:12
136.3 C 222kPa 8:42
136.4 C 221kPa 9:12
136.5 C 222kPa 9:42
136.4 C 220kPa 10:12
136.4 C 222kPa 10:42
136.4 C 223kPa 11:12
VENTING 11:12
AIR DRYING 12:17
CYCLE COMPLETE 42:17

Digital Signature #
59F7B04752E2EAA0
