
Midmark M11 - v1.0.0
Total Cycles: 3489

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2023 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.8 F 0.0
2:00 186.6 F 3.3
4:00 190.0 F 7.5
6:00 210.2 F 12.5
8:00 223.7 F 7.4
10:00 236.2 F 12.7
12:00 236.7 F 9.7
14:00 237.0 F 8.9
16:00 248.3 F 14.4
18:00 260.2 F 21.0
20:00 270.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.1
2:00 271.6 F 28.3
3:00 272.7 F 29.2
4:00 272.1 F 28.7

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 209.5 F 0.0
10:00 227.5 F 0.0
15:00 224.5 F 0.0
20:00 221.5 F 0.0
25:00 220.3 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

07/05/2023 02:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____