
Midmark M11 - v1.0.5
Total Cycles: 3776

Name: RITTER2
Sterilizer ID: V2313228
07/12/2024 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.7 F 0.0
2:00 152.1 F 1.8
4:00 167.9 F 5.1
6:00 191.9 F 9.5
8:00 215.3 F 7.2
10:00 224.9 F 9.2
12:00 233.1 F 9.9
14:00 242.1 F 12.3
16:00 240.3 F 11.2
18:00 253.8 F 17.8
20:00 265.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.3 F 29.6
2:00 271.9 F 29.9
3:00 272.1 F 29.9
4:00 272.2 F 30.3

Min 271.0 F 29.1
Max 274.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 200.7 F 0.0
10:00 190.7 F 0.0
15:00 187.0 F 0.0
20:00 182.9 F 0.0
25:00 178.9 F 0.0
30:00 177.7 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

07/12/2024 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____