
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007343
6:54 09/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:54
STERILIZING 4:00
136.4 C 220kPa 4:00
136.1 C 221kPa 4:30
136.4 C 219kPa 5:00
136.2 C 221kPa 5:30
136.1 C 220kPa 6:00
136.3 C 222kPa 6:30
136.4 C 222kPa 7:00
136.2 C 218kPa 7:30
136.2 C 218kPa 7:30
VENTING 7:30
AIR DRYING 8:18
CYCLE COMPLETE 38:18

Digital Signature #
9F663311D05BA619
