
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001664
15:29 31/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:43
STERILIZING 4:42
136.5 C 224kPa 4:42
136.4 C 226kPa 5:12
136.3 C 224kPa 5:42
136.3 C 223kPa 6:12
136.2 C 222kPa 6:42
136.1 C 223kPa 7:12
136.4 C 226kPa 7:42
136.4 C 226kPa 8:12
136.4 C 226kPa 8:12
VENTING 8:13
AIR DRYING 9:07
CYCLE COMPLETE 9:08

Digital Signature #
2134E88912BAC38C
