
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002608
14:57 25/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:08
STERILIZING 4:15
136.4 C 227kPa 4:15
136.2 C 224kPa 4:45
136.3 C 225kPa 5:15
136.4 C 223kPa 5:45
136.2 C 224kPa 6:15
136.1 C 222kPa 6:45
136.4 C 226kPa 7:15
136.1 C 222kPa 7:45
136.1 C 222kPa 7:45
VENTING 7:45
AIR DRYING 8:41
CYCLE COMPLETE 27:47

Digital Signature #
7BBDF629ECEA64A3
