
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000354
10:56 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:31
STERILIZING 3:40
136.5 C 233kPa 3:40
136.4 C 233kPa 4:10
136.4 C 231kPa 4:40
136.2 C 232kPa 5:10
136.4 C 233kPa 5:40
136.4 C 231kPa 6:10
136.3 C 229kPa 6:40
136.4 C 230kPa 7:10
136.4 C 230kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 8:10

Digital Signature #
0D984358888AF068
