
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003253
9:46 02/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:53
STERILIZING 3:56
136.4 C 220kPa 3:56
135.9 C 220kPa 4:26
136.4 C 223kPa 4:56
136.4 C 222kPa 5:26
136.2 C 223kPa 5:56
136.4 C 224kPa 6:26
136.4 C 222kPa 6:56
136.4 C 224kPa 7:26
136.4 C 224kPa 7:26
VENTING 7:26
AIR DRYING 8:18
CYCLE COMPLETE 8:23

Digital Signature #
0B4980D2CF1CEC25
