
Midmark M11 - v1.0.5
Total Cycles: 413

Name: RITTER2
Sterilizer ID: V2313228
10/04/2021 07:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 155.8 F 1.9
4:00 163.6 F 5.6
6:00 188.9 F 10.4
8:00 209.6 F 5.3
10:00 224.2 F 10.9
12:00 229.7 F 8.1
14:00 240.1 F 11.4
16:00 244.5 F 13.2
18:00 257.0 F 19.9
20:00 268.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.7 F 29.8
2:00 271.6 F 29.4
3:00 272.4 F 29.8
4:00 271.6 F 29.4

Min 271.4 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 228.8 F 0.0
10:00 210.2 F 0.0
15:00 210.1 F 0.0
20:00 204.7 F 0.0
25:00 181.8 F 0.0
30:00 182.3 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

10/04/2021 08:52 AM
ACCEPTED: X
REJECTED:
Operator: U01