

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000243  
11:03 29/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:48  
PRESSURIZING 3:49  
STERILIZING 5:12  
136.0 C 230kPa 5:12  
136.4 C 233kPa 5:42  
136.3 C 233kPa 6:12  
136.4 C 233kPa 6:42  
136.3 C 233kPa 7:12  
136.3 C 231kPa 7:42  
136.2 C 232kPa 8:12  
136.4 C 232kPa 8:42  
136.4 C 232kPa 8:42  
VENTING 8:42  
AIR DRYING 9:51  
CYCLE COMPLETE 15:47  
-----

Digital Signature #  
75D102F1F34FB0E7  
-----