
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000495
10:23 30/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:51
STERILIZING 4:17
136.4 C 231kPa 4:17
136.3 C 231kPa 4:47
136.4 C 233kPa 5:17
136.4 C 233kPa 5:47
136.4 C 230kPa 6:17
136.5 C 233kPa 6:47
136.2 C 230kPa 7:17
136.3 C 231kPa 7:47
136.3 C 231kPa 7:47
VENTING 7:47
AIR DRYING 8:37
CYCLE COMPLETE 19:51

Digital Signature #
C223DA9DEBF3B515
