

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3613

Name: RITTER2  
Sterilizer ID: V2313228  
05/14/2024 08:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 76.2 F 0.0  
4:00 79.9 F 0.0  
6:00 106.1 F 0.5  
8:00 151.4 F 3.5  
10:00 175.9 F 6.6  
12:00 193.9 F 10.5  
14:00 220.4 F 5.6  
16:00 223.2 F 6.9  
18:00 234.1 F 10.8  
20:00 230.4 F 7.6  
22:00 241.9 F 12.0  
24:00 238.5 F 10.0  
26:00 251.3 F 16.2  
28:00 262.2 F 22.5  
30:00 272.3 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.2 F 29.0  
2:00 272.8 F 29.5  
3:00 271.9 F 29.0  
4:00 272.3 F 28.9

Min 271.7 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 212.3 F 0.0  
10:00 223.4 F 0.0  
15:00 202.0 F 0.0  
20:00 205.5 F 0.0  
25:00 211.9 F 0.0  
30:00 215.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:26  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:08:24  
CYCLE COMPLETE

05/14/2024 09:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_