
Midmark M11 - v1.0.5
Total Cycles: 1177

Name: RITTER2
Sterilizer ID: V2313228
06/08/2022 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.1 F 0.0
2:00 193.1 F 3.8
4:00 207.7 F 9.1
6:00 230.6 F 8.2
8:00 233.2 F 10.9
10:00 235.1 F 8.9
12:00 239.8 F 10.1
14:00 248.4 F 14.6
16:00 260.5 F 21.5
18:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.3 F 29.5
2:00 271.5 F 28.6
3:00 272.3 F 29.5
4:00 271.8 F 28.9

Min 271.2 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 229.7 F 0.0
10:00 249.0 F 0.0
15:00 235.6 F 0.0
20:00 226.2 F 0.0
25:00 226.5 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 18:28
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

06/08/2022 04:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____