
Midmark M11 - v1.0.5
Total Cycles: 3455

Name: RITTER2
Sterilizer ID: V2313228
03/15/2024 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 74.1 F 0.0
4:00 80.0 F 0.0
6:00 109.7 F 1.1
8:00 146.2 F 3.7
10:00 172.9 F 6.5
12:00 186.2 F 10.3
14:00 205.2 F 7.2
16:00 216.9 F 6.3
18:00 227.9 F 10.1
20:00 232.6 F 9.8
22:00 237.8 F 10.8
24:00 234.8 F 8.2
26:00 246.6 F 13.9
28:00 257.0 F 19.5
30:00 266.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.6
2:00 271.5 F 28.8
3:00 272.4 F 29.7
4:00 271.9 F 29.0

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 232.5 F 0.0
10:00 196.5 F 0.0
15:00 191.8 F 0.0
20:00 188.7 F 0.0
25:00 187.6 F 0.0
30:00 189.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:55
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:10:04
CYCLE COMPLETE

03/15/2024 09:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____