
Midmark M11 - v1.0.0
Total Cycles: 3837

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2023 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 172.1 F 0.5
4:00 195.6 F 4.8
6:00 214.2 F 9.8
8:00 228.9 F 7.7
10:00 236.7 F 11.2
12:00 235.2 F 8.8
14:00 240.3 F 10.0
16:00 247.8 F 14.0
18:00 260.0 F 20.6
20:00 270.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.1 F 28.3
2:00 272.4 F 28.7
3:00 271.6 F 28.3
4:00 272.1 F 28.3

Min 271.4 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 214.7 F 0.0
10:00 224.3 F 0.0
15:00 227.1 F 0.0
20:00 227.9 F 0.0
25:00 229.4 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

09/28/2023 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____