
Midmark M11 - v1.0.0
Total Cycles: 5242

Name: RITTER 1
Sterilizer ID: V2178198
01/14/2025 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 167.7 F 0.9
4:00 171.3 F 4.1
6:00 186.7 F 7.7
8:00 203.8 F 11.8
10:00 216.2 F 4.9
12:00 226.8 F 9.2
14:00 238.1 F 13.5
16:00 237.1 F 10.1
18:00 237.1 F 8.8
20:00 246.9 F 13.8
22:00 257.9 F 19.7
24:00 267.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.2
2:00 271.8 F 28.6
3:00 271.5 F 28.7
4:00 272.4 F 29.2

Min 271.3 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 207.4 F 0.0
10:00 204.9 F 0.0
15:00 208.6 F 0.0
20:00 210.9 F 0.0
25:00 210.4 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:26
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:04:10
CYCLE COMPLETE

01/14/2025 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____