
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000511
11:33 21/05/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 3:16
CYCLE ABORTED 4:10
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:10

Digital Signature #
503601649AC1E5BF
