
Midmark M11 - v1.0.5
Total Cycles: 3211

Name: RITTER2
Sterilizer ID: V2313228
12/13/2023 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.8 F 0.0
2:00 105.2 F 0.1
4:00 119.9 F 1.3
6:00 146.2 F 3.9
8:00 168.8 F 7.1
10:00 188.9 F 10.8
12:00 210.4 F 5.2
14:00 214.4 F 8.1
16:00 227.6 F 12.5
18:00 232.6 F 9.1
20:00 238.8 F 10.9
22:00 243.4 F 12.9
24:00 255.1 F 18.8
26:00 264.6 F 25.0
28:00 271.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.9
1:00 273.1 F 30.2
2:00 272.3 F 30.4
3:00 272.4 F 30.1
4:00 272.4 F 30.6

Min 271.7 F 29.8
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 241.5 F 0.0
10:00 219.6 F 0.0
15:00 206.3 F 0.0
20:00 185.3 F 0.0
25:00 173.1 F 0.0
30:00 171.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 28:51
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:25
CYCLE COMPLETE

12/13/2023 01:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____