
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006414
11:15 05/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:26
STERILIZING 3:30
136.4 C 222kPa 3:30
136.4 C 223kPa 4:00
136.3 C 221kPa 4:30
136.1 C 221kPa 5:00
136.4 C 224kPa 5:30
136.3 C 223kPa 6:00
136.4 C 223kPa 6:30
136.4 C 224kPa 7:00
136.4 C 224kPa 7:00
VENTING 7:00
AIR DRYING 7:46
CYCLE COMPLETE 7:50

Digital Signature #
AE58DEF278B12C42
