
Midmark M11 - v1.0.5
Total Cycles: 4383

Name: RITTER2
Sterilizer ID: V2313228
02/26/2025 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 161.8 F 2.7
4:00 186.5 F 7.1
6:00 205.3 F 11.3
8:00 220.5 F 5.5
10:00 231.8 F 10.5
12:00 233.3 F 8.0
14:00 244.2 F 13.5
16:00 243.5 F 12.5
18:00 255.6 F 18.8
20:00 266.4 F 25.4
22:00 273.0 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.8
1:00 272.3 F 30.0
2:00 272.0 F 29.7
3:00 272.0 F 30.1
4:00 271.8 F 29.8

Min 271.8 F 29.7
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 204.8 F 0.0
10:00 210.4 F 0.0
15:00 211.0 F 0.0
20:00 209.9 F 0.0
25:00 209.9 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 22:16
STERILIZING: 4:00
VENTING: 4:04
DRYING: 30:00
TOTAL CYCLE: 01:02:48
CYCLE COMPLETE

02/26/2025 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____