
Midmark M11 - v1.0.0
Total Cycles: 4220

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2024 05:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.4 F 0.0
2:00 199.4 F 2.4
4:00 204.5 F 6.9
6:00 210.8 F 12.4
8:00 224.0 F 6.5
10:00 236.7 F 11.7
12:00 236.3 F 8.7
14:00 240.6 F 9.9
16:00 248.3 F 13.9
18:00 260.8 F 20.8
20:00 271.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.7
1:00 272.7 F 28.4
2:00 271.4 F 27.7
3:00 272.8 F 28.6
4:00 271.8 F 27.8

Min 271.4 F 27.6
Max 272.9 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 208.8 F 0.0
10:00 212.6 F 0.0
15:00 208.8 F 0.0
20:00 203.5 F 0.0
25:00 204.7 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

01/23/2024 06:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____