
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005669
12:48 05/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:09
STERILIZING 4:14
136.4 C 223kPa 4:14
136.2 C 223kPa 4:44
136.4 C 225kPa 5:14
136.2 C 219kPa 5:44
136.4 C 225kPa 6:14
136.3 C 223kPa 6:44
136.2 C 221kPa 7:14
136.4 C 223kPa 7:44
136.4 C 223kPa 7:44
VENTING 7:44
AIR DRYING 8:37
CYCLE COMPLETE 8:41

Digital Signature #
E570DC6B0C7D8D4B
