
Midmark M11 - v1.0.0
Total Cycles: 1055

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2021 01:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.8 F 0.0
2:00 143.6 F 0.8
4:00 152.6 F 3.2
6:00 177.6 F 6.3
8:00 199.0 F 10.5
10:00 211.6 F 5.0
12:00 223.9 F 8.1
14:00 235.0 F 12.7
16:00 235.5 F 9.3
18:00 239.1 F 9.9
20:00 246.0 F 13.4
22:00 257.2 F 19.5
24:00 267.7 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.8 F 29.3
2:00 271.7 F 28.6
3:00 272.8 F 29.5
4:00 272.2 F 28.9

Min 271.6 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 195.8 F 0.0
10:00 197.2 F 0.0
15:00 178.3 F 0.0
20:00 177.4 F 0.0
25:00 175.7 F 0.0
30:00 174.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:25
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

07/27/2021 06:33 AM
ACCEPTED: X
REJECTED:
Operator: U01