

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2148

Name: RITTER2  
Sterilizer ID: V2313228  
03/15/2023 07:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.3 F 0.0  
2:00 78.6 F 0.0  
4:00 87.4 F 0.0  
6:00 123.4 F 0.6  
8:00 154.4 F 3.5  
10:00 176.0 F 7.0  
12:00 199.6 F 10.7  
14:00 222.5 F 5.4  
16:00 222.5 F 7.4  
18:00 231.6 F 11.5  
20:00 231.8 F 7.6  
22:00 244.1 F 13.2  
24:00 242.1 F 11.7  
26:00 254.0 F 17.8  
28:00 265.1 F 24.4  
30:00 271.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.3 F 29.4  
2:00 272.5 F 29.8  
3:00 271.6 F 29.1  
4:00 272.8 F 29.6  
  
Min 271.5 F 29.0  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
5:00 222.2 F 0.0  
10:00 185.4 F 0.0  
15:00 181.9 F 0.0  
20:00 179.7 F 0.0  
25:00 178.3 F 0.0  
30:00 177.9 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:35  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:08:33  
CYCLE COMPLETE

03/15/2023 08:32 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_