
Midmark M11 - v1.0.5
Total Cycles: 859

Name: RITTER2
Sterilizer ID: V2313228
03/02/2022 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.2 F 0.0
2:00 149.1 F 2.5
4:00 171.5 F 6.1
6:00 189.4 F 10.9
8:00 213.6 F 4.9
10:00 224.6 F 10.3
12:00 229.2 F 7.5
14:00 241.4 F 13.1
16:00 241.7 F 12.0
18:00 254.8 F 18.8
20:00 266.3 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.4
1:00 272.2 F 30.4
2:00 271.9 F 30.0
3:00 272.5 F 30.1
4:00 272.7 F 30.5

Min 271.0 F 29.3
Max 273.2 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 207.6 F 0.0
10:00 219.7 F 0.0
15:00 222.2 F 0.0
20:00 219.9 F 0.0
25:00 218.4 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

03/02/2022 02:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____