
Midmark M11 - v1.0.5
Total Cycles: 420

Name: RITTER2
Sterilizer ID: V2313228
10/06/2021 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.2 F 0.0
2:00 101.1 F 0.5
4:00 102.2 F 2.0
6:00 132.9 F 4.5
8:00 158.3 F 7.9
10:00 181.4 F 11.8
12:00 213.4 F 4.4
14:00 219.2 F 8.4
16:00 229.6 F 12.6
18:00 233.6 F 8.6
20:00 242.6 F 12.1
22:00 243.1 F 12.1
24:00 254.5 F 18.0
26:00 264.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.8 F 29.3
2:00 272.1 F 29.6
3:00 271.4 F 29.2
4:00 271.4 F 28.8

Min 271.1 F 28.8
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 232.0 F 0.0
10:00 225.1 F 0.0
15:00 207.8 F 0.0
20:00 182.3 F 0.0
25:00 187.9 F 0.0
30:00 194.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 27:51
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:06:01
CYCLE COMPLETE

10/06/2021 11:33 AM
ACCEPTED: X
REJECTED:
Operator: U01