

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000313  
9:00 27/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:24  
PRESSURIZING 3:31  
STERILIZING 4:44  
136.0 C 229kPa 4:44  
136.4 C 232kPa 5:14  
136.3 C 231kPa 5:44  
136.0 C 228kPa 6:14  
136.3 C 232kPa 6:44  
136.4 C 232kPa 7:14  
136.4 C 232kPa 7:44  
136.4 C 230kPa 8:14  
136.4 C 230kPa 8:14  
VENTING 8:14  
AIR DRYING 9:09  
CYCLE COMPLETE 24:09

-----  
Digital Signature #  
59CE407757B3AFB7  
-----